


PRODUCT SPECIFICATION

<b>Product name</b>	Vinyl Examination Gloves
<b>Range</b>	Premier Protector
<b>Type</b>	Powder free, non-sterile
<b>Marking</b>	Product name, Premier Logo, Size, Code Ref., Lot Number, Expiry date, CE Mark and single use symbols
<b>Shape</b>	Standard cuff length and ambidextrous.
<b>Material</b>	Vinyl
<b>Colour</b>	Clear

Size	Length (mm) (according to EN455-2)	Palm width (mm) (according to EN455-2)	Product code
<b>S</b>	240	84 ± 3	P5850
<b>M</b>	240	94 ± 3	P5855
<b>L</b>	240	105 ± 3	P5860
<b>XL</b>	240	113 ± 3	P5865

<b>Primary use</b>	In the medical field to protect patient and user from cross-contamination when conducting medical examinations, diagnostic and therapeutic procedures and for handling contaminated medical material
--------------------	--

<b>Packaging</b>	Dispenser box (DB) S; M; L; XL.	100	items
	Outer Carton (OC)	10	DB

This product complies with Commission Regulation (EU) NO 10/2011 ( contact with food)	
---	--

<b>Residual powder</b> (according to EN 455-3)	< 2 mg/glove
---	--------------

<b>Thickness</b> (minimum at palm)	0.15 mm
---------------------------------------	---------

<b>Force at break</b> (according to EN 455-2)	≥ 3.6 N (during shelf life)
---	-----------------------------

<b>Quality certification</b>	ISO 13485
<b>Conformity to guidelines</b>	Class 1 - Medical Device Directive 93/42/EEC
<b>Conformity to standards</b>	EN 455 parts 1,2 &3
<b>Storage conditions</b>	Store in a cool dry place away from heat and direct sunlight
<b>Shelf life</b>	3 Years if stored as recommended above
<b>Additional information</b>	Powder free, latex free and with a smooth surface, the gloves offer outstanding value without compromise on quality.

*Medical Device and Reporting System:  
According to the official criteria of MDD, incidents caused by this device must be reported immediately to our offices.*

<b>REVISION:</b>	3.0	<b>DATE</b>	23/07/2015	<b>AUTHOR</b>	J Bertran
<b>NOTES:</b>	Periodic review, DSN format and Shermond logo updated.				
<b>REVISION:</b>	4.0	<b>DATE</b>	16/06/2017	<b>AUTHOR</b>	F Li
<b>NOTES:</b>	Update of food safe compliance				

The information provided within this Data Sheet is correct as of last revision but it is subject to change without notice.

Shermond T: +44 (0) 1530 278 111 W: <a href="http://www.shermond.com">www.shermond.com</a>	Unit 6, Delta Park Industrial Estate Millmarsh Lane, London Enfield, EN3 7QJ
--	--